

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082922\
 Data File : VV027557.D
 Acq On : 29 Aug 2022 12:37
 Operator : SY/MD
 Sample : VV0829WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK216

Quant Time: Aug 30 04:18:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 27 03:12:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	134275	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	142524	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	63378	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	40204	4.207	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.200%
7) Chloroethane-d5	1.568	69	43058	6.156	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	123.200%
11) 1,1-Dichloroethene-d2	2.105	63	75597	3.392	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.800%
20) 2-Butanone-d5	3.892	46	40365	33.940	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	67.880%
24) Chloroform-d	4.342	84	76543	4.006	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.200%
26) 1,2-Dichloroethane-d4	5.031	65	40702	4.924	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.400%
32) Benzene-d6	5.047	84	151686	3.842	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.800%
36) 1,2-Dichloropropane-d6	6.066	67	43925	3.827	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	76.600%
41) Toluene-d8	7.313	98	124758	3.500	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.000%
43) trans-1,3-Dichloroprop...	7.625	79	14626	3.722	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.400%
46) 2-Hexanone-d5	8.092	63	53974	41.724	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.440%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	33188	4.541	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	45688	4.863	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082922\
Data File : VV027557.D
Acq On : 29 Aug 2022 12:37
Operator : SY/MD
Sample : VV0829WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK216

Quant Time: Aug 30 04:18:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 27 03:12:26 2022
Response via : Initial Calibration

