

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082922\
 Data File : VV027572.D
 Acq On : 29 Aug 2022 19:13
 Operator : SY/MD
 Sample : N4402-06
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBVB1

Quant Time: Aug 30 04:20:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 27 03:12:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	106481	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	104800	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	41649	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	31107	4.105	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.000%
7) Chloroethane-d5	1.564	69	34516	6.223	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	124.400%
11) 1,1-Dichloroethene-d2	2.105	63	56989	3.225	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.400%
20) 2-Butanone-d5	3.892	46	60931	64.606	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	129.220%
24) Chloroform-d	4.346	84	66517	4.390	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.800%
26) 1,2-Dichloroethane-d4	5.034	65	34616	5.281	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.600%
32) Benzene-d6	5.050	84	115770	3.988	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.800%
36) 1,2-Dichloropropane-d6	6.069	67	51376	6.088	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	121.800%
41) Toluene-d8	7.317	98	95001	3.625	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.400%
43) trans-1,3-Dichloroprop...	7.628	79	8743	3.026	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	60.600%
46) 2-Hexanone-d5	8.091	63	46660	49.054	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.100%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	36162	6.729	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	134.600%#
66) 1,2-Dichlorobenzene-d4	11.622	152	30250	4.900	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.000%
Target Compounds						
17) Methyl tert-butyl Ether	2.767	73	1746	0.105	ug/L	94
25) Chloroform	4.375	83	6468	0.346	ug/L	86
47) Tetrachloroethene	7.976	164	2919	0.409	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082922\
Data File : VV027572.D
Acq On : 29 Aug 2022 19:13
Operator : SY/MD
Sample : N4402-06
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBVB1

Quant Time: Aug 30 04:20:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 27 03:12:26 2022
Response via : Initial Calibration

