


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082922\
Data File : VV027570.D
Acq On : 29 Aug 2022 18:24
Operator : SY/MD
Sample : N4402-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBVA8

Manual IntegrationsAPPROVED

Quant Time: Aug 30 04:20:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 27 03:12:26 2022
Response via : Initial Calibration

Reviewed By :Krupa Patel 08/30/2022
Supervised By :Mahesh Dadoda 08/30/2022






Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082922\
Data File : VV027570.D
Acq On : 29 Aug 2022 18:24
Operator : SY/MD
Sample : N4402-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBVA8

Manual IntegrationsAPPROVED

Reviewed By :Krupa Patel 08/30/2022
Supervised By :Mahesh Dadoda 08/30/2022

Quant Time: Aug 30 04:20:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 27 03:12:26 2022
Response via : Initial Calibration



TIC: VV027570.D\data.ms

(34) Trichloroethene (T)

5.925min (+ 0.016) 0.20 ug/L

response 2338

Ion	Exp%	Act%
95.00	100.00	100.00
96.95	64.30	63.82
131.95	93.80	98.18
129.95	99.00	100.55




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082922\
Data File : VV027570.D
Acq On : 29 Aug 2022 18:24
Operator : SY/MD
Sample : N4402-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBVA8

Manual IntegrationsAPPROVED

Quant Time: Aug 30 04:20:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 27 03:12:26 2022
Response via : Initial Calibration

Reviewed By :Krupa Patel 08/30/2022
Supervised By :Mahesh Dadoda 08/30/2022



TIC: VV027570.D\data.ms

(34) Trichloroethene (T)

5.941min (+ 0.032) 0.28 ug/L m

response 3355

Ion	Exp%	Act%
95.00	100.00	100.00
96.95	64.30	64.69
131.95	93.80	106.38
129.95	99.00	108.16

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV082922\
 Data File : VV027570.D
 Acq On : 29 Aug 2022 18:24
 Operator : SY/MD
 Sample : N4402-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBVA8

Manual IntegrationsAPPROVED

Quant Time: Aug 30 04:20:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 27 03:12:26 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 08/30/2022
 Supervised By :Mahesh Dadoda 08/30/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	138814	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	135700	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	54639	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	36659	3.711	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	74.200%	
7) Chloroethane-d5	1.565	69	42769	5.915	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	118.400%	
11) 1,1-Dichloroethene-d2	2.105	63	69784	3.029	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	60.600%	
20) 2-Butanone-d5	3.892	46	56465	45.926	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	91.860%	
24) Chloroform-d	4.346	84	88932	4.502	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.000%	
26) 1,2-Dichloroethane-d4	5.034	65	44179	5.170	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	103.400%	
32) Benzene-d6	5.047	84	165569	4.405	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.000%	
36) 1,2-Dichloropropane-d6	6.069	67	51909	4.750	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	95.000%	
41) Toluene-d8	7.317	98	124185	3.659	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	73.200%	
43) trans-1,3-Dichloroprop...	7.629	79	10144	2.711	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	54.200%#	
46) 2-Hexanone-d5	8.092	63	51233	41.597	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	83.200%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	33211	4.772	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	95.400%	
66) 1,2-Dichlorobenzene-d4	11.625	152	41438	5.116	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	102.400%	
Target Compounds						
17) Methyl tert-butyl Ether	2.767	73	16442	0.757	ug/L	# 92
25) Chloroform	4.375	83	12858	0.528	ug/L	98
34) Trichloroethene	5.941	95	3355m	0.283	ug/L	
47) Tetrachloroethene	7.973	164	32488	3.512	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed