

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV083018\
 Data File : VV007277.D
 Acq On : 30 Aug 2018 22:34
 Operator : SY/MD
 Sample : J4633-09
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Aug 31 02:53:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR083018WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 31 02:32:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	293233	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	295923	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	115179	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	125340	4.88	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.60%
7) Chloroethane-d5	1.58	69	99723	5.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.00%
11) 1,1-Dichloroethene-d2	2.13	63	160318	3.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.80%
20) 2-Butanone-d5	3.97	46	231167	46.19	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.38%
24) Chloroform-d	4.41	84	216639	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
26) 1,2-Dichloroethane-d4	5.09	65	103876	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.40%
32) Benzene-d6	5.10	84	430055	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
36) 1,2-Dichloropropane-d6	6.12	67	137608	4.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.60%
41) Toluene-d8	7.37	98	338068	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.00%
43) trans-1,3-Dichloropropene-	7.67	79	42592	4.21	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.20%
46) 2-Hexanone-d5	8.14	63	128079	37.03	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.06%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	95815	4.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	116343	5.06	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV083018\
Data File : VV007277.D
Acq On : 30 Aug 2018 22:34
Operator : SY/MD
Sample : J4633-09
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Aug 31 02:53:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR083018WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 31 02:32:21 2018
Response via : Initial Calibration

