

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV083018\
 Data File : VV007272.D
 Acq On : 30 Aug 2018 19:52
 Operator : SY/MD
 Sample : VV0830WBL01
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK68

Quant Time: Aug 31 02:37:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR083018WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 31 02:32:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	294284	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	294033	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	119392	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	118658	4.60	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.00%
7) Chloroethane-d5	1.58	69	99791	4.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.80%
11) 1,1-Dichloroethene-d2	2.13	63	156223	3.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.60%
20) 2-Butanone-d5	3.98	46	243057	48.39	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.78%
24) Chloroform-d	4.40	84	210696	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.00%
26) 1,2-Dichloroethane-d4	5.09	65	101332	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
32) Benzene-d6	5.10	84	429394	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
36) 1,2-Dichloropropane-d6	6.12	67	136405	4.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.40%
41) Toluene-d8	7.37	98	337534	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.40%
43) trans-1,3-Dichloropropene-	7.67	79	43591	4.34	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.80%
46) 2-Hexanone-d5	8.15	63	147915	43.03	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.06%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	93801	4.60	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	116927	4.90	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV083018\
Data File : VV007272.D
Acq On : 30 Aug 2018 19:52
Operator : SY/MD
Sample : VV0830WBL01
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK68

Quant Time: Aug 31 02:37:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR083018WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 31 02:32:21 2018
Response via : Initial Calibration

