

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV083019\
 Data File : VV012447.D
 Acq On : 30 Aug 2019 10:34
 Operator : SY/MD
 Sample : K4501-03
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GANL8

Quant Time: Aug 31 00:38:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR082919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 31 00:07:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	228498	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	227235	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	83120	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	11578	0.57	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	11.40%#
7) Chloroethane-d5	1.58	69	12959	0.88	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	17.60%#
11) 1,1-Dichloroethene-d2	2.13	63	28197	0.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	16.80%#
20) 2-Butanone-d5	3.93	46	164390	39.60	ug/L	-0.04
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.20%
24) Chloroform-d	4.40	84	97411	2.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	55.60%#
26) 1,2-Dichloroethane-d4	5.08	65	41712	2.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	49.00%#
32) Benzene-d6	5.10	84	128843	1.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	38.00%#
36) 1,2-Dichloropropane-d6	6.12	67	62228	2.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	57.80%#
41) Toluene-d8	7.36	98	118373	1.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	38.80%#
43) trans-1,3-Dichloropropene-	7.66	79	20877	2.74	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	54.80%#
46) 2-Hexanone-d5	8.13	63	103057	34.45	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	68.90%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	57686	3.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	75.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	68492	4.24	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	84.80%

Target Compounds					Qvalue
13) Acetone	2.19	43	50435	16.629 ug/L	95
21) 2-Butanone	4.02	43	20763	3.724 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV083019\
Data File : VV012447.D
Acq On : 30 Aug 2019 10:34
Operator : SY/MD
Sample : K4501-03
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GANL8

Quant Time: Aug 31 00:38:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR082919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 31 00:07:21 2019
Response via : Initial Calibration

