

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083021\
 Data File : VV022019.D
 Acq On : 30 Aug 2021 21:53
 Operator : SY/MD
 Sample : M3566-15
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKJ0

Quant Time: Aug 31 04:31:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 31 04:25:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	362263	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	349552	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	181764	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	91171	36.342	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.680%
7) Chloroethane-d5	1.564	69	80794	41.240	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.480%
11) 1,1-Dichloroethene-d2	2.105	63	128080	30.870	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.740%
21) 2-Butanone-d5	3.876	46	165047	95.929	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.930%
24) Chloroform-d	4.352	84	199448	43.601	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.200%
26) 1,2-Dichloroethane-d4	5.034	65	123942	46.000	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.000%
32) Benzene-d6	5.053	84	416559	43.936	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.880%
36) 1,2-Dichloropropane-d6	6.072	67	134262	45.029	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.060%
41) Toluene-d8	7.320	98	386330	43.166	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.340%
43) trans-1,3-Dichloroprop...	7.625	79	56523	39.829	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.660%
47) 2-Hexanone-d5	8.091	63	104715	91.530	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.530%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	178750	46.809	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.620%
66) 1,2-Dichlorobenzene-d4	11.628	152	165742	46.384	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.760%
Target Compounds						
					Qvalue	
25) Chloroform	4.378	83	5269	1.106	ug/L	88
63) 1,2,4-Trimethylbenzene	10.921	105	8665	0.889	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083021\
Data File : VV022019.D
Acq On : 30 Aug 2021 21:53
Operator : SY/MD
Sample : M3566-15
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKJ0

Quant Time: Aug 31 04:31:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 31 04:25:57 2021
Response via : Initial Calibration

