

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083021\
 Data File : VV022024.D
 Acq On : 30 Aug 2021 23:54
 Operator : SY/MD
 Sample : M3566-20
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKJ5

Quant Time: Aug 31 04:32:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 31 04:25:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	361795	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	347487	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	174203	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	90514	36.126	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	72.260%		
7) Chloroethane-d5	1.564	69	79845	40.808	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	81.620%		
11) 1,1-Dichloroethene-d2	2.105	63	128466	31.003	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	62.000%		
21) 2-Butanone-d5	3.876	46	165131	96.101	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	96.100%		
24) Chloroform-d	4.352	84	199781	43.730	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	87.460%		
26) 1,2-Dichloroethane-d4	5.034	65	125098	46.489	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.980%		
32) Benzene-d6	5.050	84	418563	44.410	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	88.820%		
36) 1,2-Dichloropropane-d6	6.072	67	133314	44.977	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	89.960%		
41) Toluene-d8	7.320	98	385901	43.374	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	86.740%		
43) trans-1,3-Dichloroprop...	7.625	79	54617	38.714	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	77.420%		
47) 2-Hexanone-d5	8.088	63	103679	91.163	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	91.160%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	180323	47.502	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	95.000%		
66) 1,2-Dichlorobenzene-d4	11.628	152	158943	46.412	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	92.820%		
Target Compounds					Qvalue	
25) Chloroform	4.378	83	6376	1.340	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083021\
Data File : VV022024.D
Acq On : 30 Aug 2021 23:54
Operator : SY/MD
Sample : M3566-20
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKJ5

Quant Time: Aug 31 04:32:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 31 04:25:57 2021
Response via : Initial Calibration

