

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083022\
 Data File : VV027584.D
 Acq On : 30 Aug 2022 14:24
 Operator : SY/MD
 Sample : N4368-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5F06

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 08/31/2022
 Supervised By :Mahesh Dadoda 08/31/2022

Quant Time: Aug 31 04:48:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Aug 31 04:45:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	150782	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	149645	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	59617	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	47070	4.386	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	87.800%	
7) Chloroethane-d5	1.564	69	40147	5.112	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	102.200%	
11) 1,1-Dichloroethene-d2	2.105	63	69960m	2.796	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	56.000%#	
20) 2-Butanone-d5	3.899	46	66548	49.830	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	99.660%	
24) Chloroform-d	4.349	84	99031	4.615	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.400%	
26) 1,2-Dichloroethane-d4	5.034	65	45575	4.910	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.200%	
32) Benzene-d6	5.047	84	179975	4.342	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.800%	
36) 1,2-Dichloropropane-d6	6.069	67	60801	5.045	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	101.000%	
41) Toluene-d8	7.317	98	134692	3.599	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	72.000%	
43) trans-1,3-Dichloroprop...	7.625	79	16180	3.921	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	78.400%	
46) 2-Hexanone-d5	8.092	63	64869	47.760	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	95.520%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	39919	5.202	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	104.000%	
66) 1,2-Dichlorobenzene-d4	11.622	152	44100	4.990	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.800%	
Target Compounds					Qvalue	
30) Cyclohexane	4.680	56	9150	0.519	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083022\
Data File : VV027584.D
Acq On : 30 Aug 2022 14:24
Operator : SY/MD
Sample : N4368-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1


Instrument :
MSVOA_V
ClientSampleId :
F5F06

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 08/31/2022
Supervised By :Mahesh Dadoda 08/31/2022


Quant Time: Aug 31 04:48:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Aug 31 04:45:39 2022
Response via : Initial Calibration





#30
Cyclohexane
Concen: 0.519 ug/L
RT: 4.680 min Scan# 1132
Delta R.T. 0.007 min
Lab File: VV027584.D
Acq: 30 Aug 2022 14:24

Instrument :
MSVOA_V
ClientSampleId :
F5F06



Tgt Ion: 56 Resp: 9150

Ion	Ratio	Lower	Upper
56	100		
69	32.1	25.4	38.0
84	92.8	69.7	104.5

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 08/31/2022
Supervised By :Mahesh Dadoda 08/31/2022

