

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV083118\
 Data File : VV007308.D
 Acq On : 31 Aug 2018 18:18
 Operator : SY/MD
 Sample : J4665-17
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0AB3

Quant Time: Sep 01 00:03:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM083118WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 31 23:57:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	605901	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	574208	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	212703	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	167510	45.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.50%
7) Chloroethane-d5	1.58	69	112357	48.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.46%
11) 1,1-Dichloroethene-d2	2.13	63	235820	34.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.38%
21) 2-Butanone-d5	3.96	46	261949	88.46	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.46%
24) Chloroform-d	4.41	84	357458	45.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.50%
26) 1,2-Dichloroethane-d4	5.09	65	225495	47.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.02%
32) Benzene-d6	5.11	84	733870	47.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.12%
36) 1,2-Dichloropropane-d6	6.12	67	246217	48.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.24%
41) Toluene-d8	7.37	98	603986	43.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.54%
43) trans-1,3-Dichloropropene-	7.67	79	98367	43.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.22%
47) 2-Hexanone-d5	8.14	63	168011	88.14	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.14%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	313109	44.93	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.86%
64) 1,2-Dichlorobenzene-d4	11.68	152	225396	52.74	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.48%


Target Compounds					Qvalue
13) Acetone	2.21	43	13665	5.353 ug/L	97
59) 1,2,3-Trichloropropane	11.30	75	24587	4.413 ug/L	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV083118\
Data File : VV007308.D
Acq On : 31 Aug 2018 18:18
Operator : SY/MD
Sample : J4665-17
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0AB3

Quant Time: Sep 01 00:03:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM083118WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 31 23:57:34 2018
Response via : Initial Calibration





#13

Acetone

Concen: 5.353 ug/L

RT: 2.21 min Scan# 409

Delta R.T. 0.00 min

Lab File: VV007308.D

Acq: 31 Aug 2018 18:18

Instrument :

MSVOA_V

ClientSampleId :


E0AB3

Tgt Ion: 43 Resp: 13665

Ion Ratio Lower Upper




43 100

58 38.0 0.0 72.2



Abundance Ion 43.00 (42.70 to 43.70): VV007291.D (-397) (-)

Ion 58.00 (57.70 to 58.70): VV007291.D (-397) (-)



#59

1,2,3-Trichloropropane

Concen: 4.413 ug/L

RT: 11.30 min Scan# 3237

Delta R.T. 0.97 min

Lab File: VV007308.D


Acq: 31 Aug 2018 18:18

Tgt Ion: 75 Resp: 24587

Ion Ratio Lower Upper

75 100

77 36.5 25.9 38.9



Abundance Ion 75.00 (74.70 to 75.70): VV007291.D (-2922) (-)

Ion 77.00 (76.70 to 77.70): VV007291.D (-2922) (-)

