

Data File : VV007297.D
Acq On : 31 Aug 2018 13:27
Operator : SY/MD
Sample : VV0831WBL01
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK70

Quant Time: Sep 01 00:13:30 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM083118WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 31 23:57:34 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	622446	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	588995	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	233152	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	181887	48.35	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.70%
7) Chloroethane-d5	1.58	69	151509	63.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	127.94%
11) 1,1-Dichloroethene-d2	2.13	63	252608	36.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.34%
21) 2-Butanone-d5	3.96	46	293063	96.34	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.34%
24) Chloroform-d	4.41	84	376697	46.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.84%
26) 1,2-Dichloroethane-d4	5.09	65	235091	48.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.44%
32) Benzene-d6	5.10	84	778895	49.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.42%
36) 1,2-Dichloropropane-d6	6.12	67	257124	48.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.98%
41) Toluene-d8	7.36	98	646913	45.70	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.40%
43) trans-1,3-Dichloropropene-	7.67	79	108273	46.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.52%
47) 2-Hexanone-d5	8.14	63	176447	90.24	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.24%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	339631	47.51	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.02%
64) 1,2-Dichlorobenzene-d4	11.68	152	243274	51.93	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.86%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV007297.D
Acq On : 31 Aug 2018 13:27
Operator : SY/MD
Sample : VV0831WBL01
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK70

Quant Time: Sep 01 00:13:30 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM083118WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 31 23:57:34 2018
Response via : Initial Calibration

