

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV083118\
 Data File : VV007298.D
 Acq On : 31 Aug 2018 14:09
 Operator : SY/MD
 Sample : J4665-11
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0AA5

Quant Time: Sep 01 00:14:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM083118WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 31 23:57:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	575582	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	554383	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	220168	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	203045	58.37	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	116.74%
7) Chloroethane-d5	1.58	69	144010	65.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	131.50%#
11) 1,1-Dichloroethene-d2	2.13	63	284453	44.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.10%
21) 2-Butanone-d5	3.96	46	332917	118.35	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	118.35%
24) Chloroform-d	4.40	84	422623	56.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.64%
26) 1,2-Dichloroethane-d4	5.09	65	264621	58.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	117.38%
32) Benzene-d6	5.10	84	872651	58.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	117.16%
36) 1,2-Dichloropropane-d6	6.12	67	293096	59.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	118.66%
41) Toluene-d8	7.36	98	744275	55.87	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.74%
43) trans-1,3-Dichloropropene-	7.67	79	122857	55.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	111.52%
47) 2-Hexanone-d5	8.15	63	210730	114.50	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	114.50%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	381789	56.74	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	113.48%
64) 1,2-Dichlorobenzene-d4	11.68	152	293597	66.37	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	132.74%#

Target Compounds

						Ovalue
5) Vinyl chloride	1.32	62	75522	16.19	ug/L	95
17) trans-1,2-Dichloroethene	2.79	96	18177	4.83	ug/L	98
19) 1,1-Dichloroethane	3.23	63	59032	7.76	ug/L	99
20) cis-1,2-Dichloroethene	3.96	96	463528	107.02	ug/L	98
30) 1,1,1-Trichloroethane	4.66	97	64384	10.20	ug/L	100
33) Benzene	5.15	78	24353	1.45	ug/L	100
34) Trichloroethene	5.96	95	1107280	272.16	ug/L	99
46) Tetrachloroethene	8.02	164	502129	147.22	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV083118\
Data File : VV007298.D
Acq On : 31 Aug 2018 14:09
Operator : SY/MD
Sample : J4665-11
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0AA5

Quant Time: Sep 01 00:14:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM083118WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 31 23:57:34 2018
Response via : Initial Calibration

