

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV083118\
 Data File : VV007332.D
 Acq On : 01 Sep 2018 04:32
 Operator : SY/MD
 Sample : J4695-18 50X
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BEGB2

Quant Time: Sep 01 05:55:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM083118WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 01 01:00:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	620366	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	592304	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	216580	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	187791	50.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	100.18%
7) Chloroethane-d5	1.58	69	124245	52.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.26%
11) 1,1-Dichloroethene-d2	2.13	63	260212	37.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.78%
21) 2-Butanone-d5	3.97	46	268379	88.52	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.52%
24) Chloroform-d	4.41	84	375681	46.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.90%
26) 1,2-Dichloroethane-d4	5.09	65	236022	48.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.14%
32) Benzene-d6	5.11	84	764657	48.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.08%
36) 1,2-Dichloropropane-d6	6.12	67	254971	48.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.62%
41) Toluene-d8	7.37	98	628229	44.14	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.28%
43) trans-1,3-Dichloropropene-	7.67	79	96201	40.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.74%
47) 2-Hexanone-d5	8.15	63	171097	87.01	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.01%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	322956	44.92	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.84%
64) 1,2-Dichlorobenzene-d4	11.68	152	221536	50.91	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.82%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV083118\
Data File : VV007332.D
Acq On : 01 Sep 2018 04:32
Operator : SY/MD
Sample : J4695-18 50X
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BEGB2

Quant Time: Sep 01 05:55:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM083118WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 01 01:00:35 2018
Response via : Initial Calibration

