

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083122\
 Data File : VV027613.D
 Acq On : 31 Aug 2022 11:43
 Operator : SY/MD
 Sample : VV0831WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK219

Quant Time: Sep 01 02:34:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 01 02:33:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	140664	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	142935	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	58606	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	46839	4.679	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.600%
7) Chloroethane-d5	1.564	69	40499	5.528	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.600%
11) 1,1-Dichloroethene-d2	2.105	63	72274	3.096	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.000%
20) 2-Butanone-d5	3.905	46	61993	49.759	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.520%
24) Chloroform-d	4.346	84	101875	5.089	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.800%
26) 1,2-Dichloroethane-d4	5.034	65	47777	5.517	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.400%
32) Benzene-d6	5.050	84	180923	4.570	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.400%
36) 1,2-Dichloropropane-d6	6.069	67	58801	5.108	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.200%
41) Toluene-d8	7.313	98	135427	3.788	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.800%
43) trans-1,3-Dichloroprop...	7.625	79	17148	4.351	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.000%
46) 2-Hexanone-d5	8.092	63	62644	48.287	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.580%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	37121	5.064	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.200%
66) 1,2-Dichlorobenzene-d4	11.625	152	44554	5.128	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.600%
Target Compounds						
16) Methylene chloride	2.507	84	3813	0.238	ug/L	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083122\
Data File : VV027613.D
Acq On : 31 Aug 2022 11:43
Operator : SY/MD
Sample : VV0831WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK219

Quant Time: Sep 01 02:34:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 01 02:33:10 2022
Response via : Initial Calibration

