

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083122\
 Data File : VV027615.D
 Acq On : 31 Aug 2022 12:30
 Operator : SY/MD
 Sample : N4368-16
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Sep 01 02:35:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 01 02:33:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	144167	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	144333	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	58397	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	44963	4.382	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.600%
7) Chloroethane-d5	1.568	69	39331	5.238	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.800%
11) 1,1-Dichloroethene-d2	2.108	63	68380	2.858	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.200%#
20) 2-Butanone-d5	3.899	46	58289	45.649	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.300%
24) Chloroform-d	4.346	84	95030	4.632	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.600%
26) 1,2-Dichloroethane-d4	5.034	65	46971	5.292	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.800%
32) Benzene-d6	5.050	84	167709	4.195	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.000%
36) 1,2-Dichloropropane-d6	6.069	67	56023	4.820	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.400%
41) Toluene-d8	7.317	98	125427	3.475	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.400%#
43) trans-1,3-Dichloroprop...	7.625	79	15643	3.931	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.600%
46) 2-Hexanone-d5	8.091	63	63820	48.717	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.440%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	36996	4.998	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	42343	4.891	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.800%
Target Compounds						Qvalue
16) Methylene chloride	2.510	84	3248	0.198	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083122\
Data File : VV027615.D
Acq On : 31 Aug 2022 12:30
Operator : SY/MD
Sample : N4368-16
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Sep 01 02:35:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR081922WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 01 02:33:10 2022
Response via : Initial Calibration

