

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW083122\
 Data File : VW027635.D
 Acq On : 31 Aug 2022 23:29
 Operator : SY/MD
 Sample : N4366-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5E72

Quant Time: Sep 01 04:15:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 01 02:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	134931	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	125373	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	48955	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.297	65	73827	5.036	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.800%
7) Chloroethane-d5	1.558	69	60508	4.934	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.600%
11) 1,1-Dichloroethene-d2	2.098	63	112454	3.920	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.400%
20) 2-Butanone-d5	3.957	46	147475	59.505	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	119.020%
24) Chloroform-d	4.339	84	139079	4.726	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
26) 1,2-Dichloroethane-d4	5.024	65	74087	5.620	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.400%
32) Benzene-d6	5.030	84	299668	5.835	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	116.800%
36) 1,2-Dichloropropane-d6	6.069	67	84566	5.324	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.400%
41) Toluene-d8	7.326	98	203928	4.585	ug/L	0.01
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.800%
43) trans-1,3-Dichloroprop...	7.635	79	23134	4.727	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.600%
46) 2-Hexanone-d5	8.107	63	81384	44.225	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.440%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	47701	4.666	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	58561	5.343	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083122\
Data File : VV027635.D
Acq On : 31 Aug 2022 23:29
Operator : SY/MD
Sample : N4366-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5E72

Quant Time: Sep 01 04:15:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 01 02:22:45 2022
Response via : Initial Calibration

