

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW083122\
 Data File : VW027653.D
 Acq On : 01 Sep 2022 06:49
 Operator : SY/MD
 Sample : N4400-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBVA7

Quant Time: Sep 01 07:07:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 01 02:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	131435	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	130902	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	54458	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	70759	4.956	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.200%
7) Chloroethane-d5	1.568	69	59479	4.979	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.600%
11) 1,1-Dichloroethene-d2	2.108	63	107177	3.835	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.800%
20) 2-Butanone-d5	3.889	46	74488	30.855	ug/L	-0.05
Spiked Amount	50.000	Range	40 - 130	Recovery	=	61.700%
24) Chloroform-d	4.346	84	135443	4.725	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
26) 1,2-Dichloroethane-d4	5.030	65	63175	4.920	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.400%
32) Benzene-d6	5.047	84	251241	4.686	ug/L	0.01
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.800%
36) 1,2-Dichloropropane-d6	6.069	67	76891	4.636	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.800%
41) Toluene-d8	7.313	98	195307	4.206	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.200%
43) trans-1,3-Dichloroprop...	7.622	79	21298	4.168	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.400%
46) 2-Hexanone-d5	8.088	63	78304	40.754	ug/L	-0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.500%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	46465	4.354	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	63032	5.170	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.400%
Target Compounds						Qvalue
17) Methyl tert-butyl Ether	2.770	73	13985	0.597	ug/L	# 95
25) Chloroform	4.375	83	21651	0.847	ug/L	94
47) Tetrachloroethene	7.976	164	5531	0.544	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083122\
Data File : VV027653.D
Acq On : 01 Sep 2022 06:49
Operator : SY/MD
Sample : N4400-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBVA7

Quant Time: Sep 01 07:07:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 01 02:22:45 2022
Response via : Initial Calibration

