

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW083122\
 Data File : VW027670.D
 Acq On : 01 Sep 2022 15:03
 Operator : SY/MD
 Sample : N4369-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5F40

Quant Time: Sep 01 23:18:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 01 07:36:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	124712	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	122880	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	50519	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	54900	4.052	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.000%
7) Chloroethane-d5	1.561	69	49178	4.338	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.800%
11) 1,1-Dichloroethene-d2	2.098	63	90152	3.400	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.000%
20) 2-Butanone-d5	3.905	46	124563	54.379	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.760%
24) Chloroform-d	4.342	84	133392	4.904	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.000%
26) 1,2-Dichloroethane-d4	5.031	65	64321	5.279	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.600%
32) Benzene-d6	5.047	84	234017	4.649	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
36) 1,2-Dichloropropane-d6	6.066	67	74315	4.773	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.400%
41) Toluene-d8	7.313	98	174912	4.013	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.200%
43) trans-1,3-Dichloroprop...	7.625	79	19530	4.072	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.400%
46) 2-Hexanone-d5	8.088	63	88572	49.107	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.220%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	50498	5.040	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	60896	5.384	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.600%
Target Compounds						
17) Methyl tert-butyl Ether	2.770	73	23129	1.040	ug/L	99
22) cis-1,2-Dichloroethene	3.905	96	3795	0.322	ug/L	81
34) Trichloroethene	5.915	95	16065	1.364	ug/L	96
47) Tetrachloroethene	7.973	164	33084	3.463	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083122\
Data File : VV027670.D
Acq On : 01 Sep 2022 15:03
Operator : SY/MD
Sample : N4369-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 62 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5F40

Quant Time: Sep 01 23:18:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 01 07:36:37 2022
Response via : Initial Calibration

