

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW083122\
 Data File : VW027658.D
 Acq On : 01 Sep 2022 10:14
 Operator : SY/MD
 Sample : N4400-12
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBVG5

Quant Time: Sep 01 23:15:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 01 07:36:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	118895	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	117616	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	49230	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	65200	5.048	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.000%
7) Chloroethane-d5	1.564	69	58521	5.415	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.400%
11) 1,1-Dichloroethene-d2	2.105	63	101106	3.999	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.000%
20) 2-Butanone-d5	3.889	46	66469	30.437	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	60.880%
24) Chloroform-d	4.342	84	126316	4.871	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.400%
26) 1,2-Dichloroethane-d4	5.031	65	58529	5.039	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.800%
32) Benzene-d6	5.047	84	233964	4.856	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
36) 1,2-Dichloropropane-d6	6.066	67	71564	4.802	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.000%
41) Toluene-d8	7.313	98	182815	4.382	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.600%
43) trans-1,3-Dichloroprop...	7.625	79	18835	4.103	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.000%
46) 2-Hexanone-d5	8.088	63	72511	42.002	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.000%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	42279	4.409	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	57303	5.199	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083122\
Data File : VV027658.D
Acq On : 01 Sep 2022 10:14
Operator : SY/MD
Sample : N4400-12
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBVG5

Quant Time: Sep 01 23:15:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 01 07:36:37 2022
Response via : Initial Calibration

