

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW083123\
 Data File : VW031959.D
 Acq On : 31 Aug 2023 14:23
 Operator : SY/MD
 Sample : 04175-22ME
 Misc : 5.77g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EZYW5ME

Quant Time: Sep 01 02:09:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 01 00:57:38 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 09/01/2023
 Supervised By : Mahesh Dadoda 09/01/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	234441	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	228961	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	119388	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	55665	37.939	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.880%
7) Chloroethane-d5	1.529	69	40627	31.587	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	63.180%#
11) 1,1-Dichloroethene-d2	2.040	65	27890	40.595	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.200%
21) 2-Butanone-d5	3.815	46	140696	121.924	ug/L	0.03
Spiked Amount	100.000	Range	40 - 130	Recovery	=	121.920%
24) Chloroform-d	4.253	84	159582	49.761	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.520%
26) 1,2-Dichloroethane-d4	4.944	65	111075	52.191	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.380%
32) Benzene-d6	4.960	84	319265	51.101	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.200%
36) 1,2-Dichloropropane-d6	5.992	67	113875	56.628	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	113.260%
41) Toluene-d8	7.246	98	282928	49.148	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.300%
43) trans-1,3-Dichloroprop...	7.558	79	55185	52.947	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	105.900%
47) 2-Hexanone-d5	8.069	63	68740m	111.529	ug/L	0.04
Spiked Amount	100.000	Range	45 - 130	Recovery	=	111.530%
56) 1,1,2,2-Tetrachloroeth...	10.165	84	155902	59.995	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	120.000%
66) 1,2-Dichlorobenzene-d4	11.571	152	118458	54.429	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.860%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV083123\
Data File : VV031959.D
Acq On : 31 Aug 2023 14:23
Operator : SY/MD
Sample : 04175-22ME
Misc : 5.77g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EZYW5ME

Quant Time: Sep 01 02:09:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 01 00:57:38 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/01/2023
Supervised By :Mahesh Dadoda 09/01/2023

