

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090119\
 Data File : VV012513.D
 Acq On : 01 Sep 2019 16:37
 Operator : SY/MD
 Sample : K4605-06
 Misc : 5mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK02

Quant Time: Sep 02 05:13:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 02 04:16:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	315301	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	302924	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	117808	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	100042	47.97	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.94%
7) Chloroethane-d5	1.58	69	109024	48.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.14%
11) 1,1-Dichloroethene-d2	2.13	63	191646	33.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.54%
21) 2-Butanone-d5	3.94	46	179975	102.56	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.56%
24) Chloroform-d	4.40	84	203795	47.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.16%
26) 1,2-Dichloroethane-d4	5.08	65	139680	50.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.32%
32) Benzene-d6	5.10	84	366734	51.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.26%
36) 1,2-Dichloropropane-d6	6.12	67	131235	50.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.08%
41) Toluene-d8	7.36	98	322127	48.02	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.04%
43) trans-1,3-Dichloropropene-	7.66	79	54745	47.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.16%
47) 2-Hexanone-d5	8.14	63	90814	89.09	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.09%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	174093	47.59	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.18%
64) 1,2-Dichlorobenzene-d4	11.67	152	124245	54.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.84%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090119\
Data File : VV012513.D
Acq On : 01 Sep 2019 16:37
Operator : SY/MD
Sample : K4605-06
Misc : 5mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK02

Quant Time: Sep 02 05:13:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090119WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 02 04:16:56 2019
Response via : Initial Calibration

