

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090219\
 Data File : VV012517.D
 Acq On : 02 Sep 2019 09:46
 Operator : SY/MD
 Sample : VSTD00131
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD00131

Manual Integrations
 APPROVED

MMDadoda
 9/3/2019 9:10:30 AM

Quant Time: Sep 03 01:57:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 03 01:44:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	167411	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	162374	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	76191	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	16225	1.07	ug/L	0.00
7) Chloroethane-d5	1.58	69	15707	1.34	ug/L	0.00
11) 1,1-Dichloroethene-d2	2.14	63	42217	1.53	ug/L	0.00
20) 2-Butanone-d5	3.99	46	27913m	9.16	ug/L	0.02
24) Chloroform-d	4.40	84	25081	0.99	ug/L	0.00
26) 1,2-Dichloroethane-d4	5.08	65	13000	1.04	ug/L	0.00
32) Benzene-d6	5.10	84	44732	0.94	ug/L	0.00
36) 1,2-Dichloropropane-d6	6.12	67	14130	0.93	ug/L	0.00
41) Toluene-d8	7.36	98	39994	0.94	ug/L	0.00
43) trans-1,3-Dichloropropene-	7.67	79	5129	0.96	ug/L	0.00
46) 2-Hexanone-d5	8.14	63	15154	7.22	ug/L	0.00
57) 1,1,2,2-Tetrachloroethane-	10.26	84	9831	0.92	ug/L	0.00
64) 1,2-Dichlorobenzene-d4	11.67	152	13946	0.95	ug/L	0.00

Target Compounds

					Qvalue
2) Dichlorodifluoromethane	1.14	85	25043	1.141 ug/L	99
3) Chloromethane	1.25	50	29324	1.279 ug/L	97
5) Vinyl chloride	1.32	62	31755	1.423 ug/L	99
6) Bromomethane	1.54	94	17761	1.433 ug/L	98
8) Chloroethane	1.60	64	20927	1.651 ug/L	96
9) Trichlorofluoromethane	1.77	101	47178	1.568 ug/L	96
10) 1,1,2-Trichloro-1,2,2-trif	2.14	101	27479	1.780 ug/L	98
12) 1,1-Dichloroethene	2.14	96	25057	1.714 ug/L	97
13) Acetone	2.23	43	49877	18.743 ug/L	96
14) Carbon disulfide	2.32	76	81341	1.814 ug/L	100
15) Methyl Acetate	2.47	43	12548	1.896 ug/L	94
16) Methylene chloride	2.53	84	28749	1.715 ug/L	99
17) Methyl tert-butyl Ether	2.81	73	40597	1.154 ug/L	98
18) trans-1,2-Dichloroethene	2.79	96	19305	1.177 ug/L	97
19) 1,1-Dichloroethane	3.23	63	37874	1.184 ug/L	99
21) 2-Butanone	4.05	43	43296	10.586 ug/L #	67
22) cis-1,2-Dichloroethene	3.96	96	19979	1.174 ug/L	88
23) Bromochloromethane	4.30	128	8116	1.113 ug/L	90
25) Chloroform	4.43	83	39979	1.194 ug/L	93
27) 1,2-Dichloroethane	5.18	62	22284	1.169 ug/L	100
29) 1,1,1-Trichloroethane	4.65	97	30675	1.176 ug/L	100
30) Cyclohexane	4.72	56	29491	1.110 ug/L	99
31) Carbon tetrachloride	4.87	117	26912	1.173 ug/L	99
33) Benzene	5.14	78	73639	1.098 ug/L	100
34) Trichloroethene	5.96	95	21034	1.225 ug/L	98
35) Methylcyclohexane	6.17	83	29605	1.134 ug/L	98
37) 1,2-Dichloropropane	6.22	63	21091	1.215 ug/L #	94
38) Bromodichloromethane	6.55	83	24986	1.145 ug/L	99
39) cis-1,3-Dichloropropene	7.07	75	23544	1.040 ug/L	99
40) 4-Methyl-2-pentanone	7.28	43	105786	10.567 ug/L	97

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090219\
 Data File : VV012517.D
 Acq On : 02 Sep 2019 09:46
 Operator : SY/MD
 Sample : VSTD00131
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD00131

Manual Integrations
 APPROVED

MMDadoda
 9/3/2019 9:10:30 AM

Quant Time: Sep 03 01:57:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 03 01:44:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) Toluene	7.43	91	78955	1.147	ug/L	99
44) trans-1,3-Dichloropropene	7.69	75	18401	1.042	ug/L	99
45) 1,1,2-Trichloroethane	7.88	97	12715	1.097	ug/L	97
47) Tetrachloroethene	8.02	164	15251	1.127	ug/L	97
48) 2-Hexanone	8.19	43	72956	10.691	ug/L	95
49) Dibromochloromethane	8.29	129	14219	1.080	ug/L	99
50) 1,2-Dibromoethane	8.40	107	10789	1.059	ug/L	99
51) Chlorobenzene	8.92	112	50566	1.160	ug/L	97
52) Ethylbenzene	9.05	91	81363	1.111	ug/L	98
53) m,p-xylene	9.18	106	29042	1.094	ug/L	93
54) o-xylene	9.59	106	26436	1.016	ug/L	98
55) Styrene	9.60	104	44606	1.026	ug/L	97
56) Isopropylbenzene	9.97	105	74930	1.087	ug/L	98
58) 1,1,2,2-Tetrachloroethane	10.28	83	13469	0.998	ug/L	97
59) 1,2,3-Trichloropropane	10.32	75	11333	1.165	ug/L	99
61) Bromoform	9.77	173	6823	0.991	ug/L	100
62) 1,3-Dichlorobenzene	11.22	146	36352	1.143	ug/L	98
63) 1,4-Dichlorobenzene	11.31	146	38151	1.195	ug/L	96
65) 1,2-Dichlorobenzene	11.69	146	33003	1.104	ug/L	97
66) 1,2-Dibromo-3-chloropropan	12.48	75	2117	1.049	ug/L #	87
67) 1,3,5-Trichlorobenzene	12.69	180	25656	1.082	ug/L	97
68) 1,2,4-trichlorobenzene	13.31	180	17693	1.018	ug/L	96
69) Naphthalene	13.55	128	21095	0.841	ug/L	98
70) 1,2,3-Trichlorobenzene	13.79	180	15457	0.942	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

