

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090219\
 Data File : VV012526.D
 Acq On : 02 Sep 2019 14:03
 Operator : SY/MD
 Sample : K4607-12
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFD68

Quant Time: Sep 03 06:29:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 03 05:44:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	150918	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	149733	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	63765	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	67033	4.57	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.40%
7) Chloroethane-d5	1.59	69	68048	4.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.00%
11) 1,1-Dichloroethene-d2	2.13	63	111029	3.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.00%
20) 2-Butanone-d5	3.96	46	131199	46.97	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.94%
24) Chloroform-d	4.40	84	103783	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.00%
26) 1,2-Dichloroethane-d4	5.08	65	56807	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.20%
32) Benzene-d6	5.10	84	200568	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
36) 1,2-Dichloropropane-d6	6.12	67	64767	4.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.80%
41) Toluene-d8	7.36	98	164649	4.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.40%
43) trans-1,3-Dichloropropene-	7.67	79	21496	4.44	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.80%
46) 2-Hexanone-d5	8.14	63	79866	44.04	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.08%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	40896	4.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	56195	4.80	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.00%

Target Compounds					Ovalue
13) Acetone	2.21	43	15309	3.863 ug/L	95
42) Toluene	7.43	91	59386	0.889 ug/L	96
55) Styrene	9.60	104	67993	1.602 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090219\
Data File : VV012526.D
Acq On : 02 Sep 2019 14:03
Operator : SY/MD
Sample : K4607-12
Misc : 25mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFD68

Quant Time: Sep 03 06:29:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 03 05:44:01 2019
Response via : Initial Calibration

