

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090219\
 Data File : VV012536.D
 Acq On : 02 Sep 2019 18:43
 Operator : SY/MD
 Sample : K4607-09
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFD65

Quant Time: Sep 03 07:09:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 03 05:44:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	150971	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	146826	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	60015	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	68397	4.66	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.20%
7) Chloroethane-d5	1.58	69	66468	4.59	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.80%
11) 1,1-Dichloroethene-d2	2.13	63	145860	4.07	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.40%
20) 2-Butanone-d5	3.96	46	140969	50.45	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.90%
24) Chloroform-d	4.40	84	106158	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
26) 1,2-Dichloroethane-d4	5.08	65	56360	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.40%
32) Benzene-d6	5.10	84	200671	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
36) 1,2-Dichloropropane-d6	6.12	67	65919	5.07	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.40%
41) Toluene-d8	7.36	98	171855	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.00%
43) trans-1,3-Dichloropropene-	7.67	79	21119	4.45	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.00%
46) 2-Hexanone-d5	8.14	63	82938	46.64	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.28%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	43988	4.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	56245	5.11	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.20%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.14	96	57652	2.952 ug/L #	77
13) Acetone	2.22	43	12209	3.079 ug/L	95
14) Carbon disulfide	2.32	76	27311	0.435 ug/L #	93
19) 1,1-Dichloroethane	3.23	63	6528	0.217 ug/L	97
22) cis-1,2-Dichloroethene	3.96	96	81477	5.049 ug/L	99
29) 1,1,1-Trichloroethane	4.66	97	12795	0.526 ug/L	98
34) Trichloroethene	5.96	95	516085	30.897 ug/L	99
47) Tetrachloroethene	8.02	164	539345	44.686 ug/L	96
70) 1,2,3-Trichlorobenzene	13.79	180	2543	0.204 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090219\
Data File : VV012536.D
Acq On : 02 Sep 2019 18:43
Operator : SY/MD
Sample : K4607-09
Misc : 25mL/MSVOA V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFD65

Quant Time: Sep 03 07:09:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 03 05:44:01 2019
Response via : Initial Calibration

