

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\V090221\
 Data File : V022080.D
 Acq On : 02 Sep 2021 16:59
 Operator : SY/MD
 Sample : M3608-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKK3

Quant Time: Sep 03 01:02:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 03 00:59:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	387593	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	354137	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	174942	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	106057	39.513	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.020%
7) Chloroethane-d5	1.561	69	83116	39.653	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.300%
11) 1,1-Dichloroethene-d2	2.105	63	149051	33.577	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.160%
21) 2-Butanone-d5	3.880	46	173903	94.470	ug/L	-0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.470%
24) Chloroform-d	4.352	84	210540	43.018	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.040%
26) 1,2-Dichloroethane-d4	5.034	65	131981	45.782	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.560%
32) Benzene-d6	5.050	84	442870	46.106	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.220%
36) 1,2-Dichloropropane-d6	6.072	67	138707	45.918	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.840%
41) Toluene-d8	7.317	98	407040	44.891	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.780%
43) trans-1,3-Dichloroprop...	7.622	79	62205	43.265	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.540%
47) 2-Hexanone-d5	8.088	63	118916	102.597	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.600%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	185872	48.044	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.080%
66) 1,2-Dichlorobenzene-d4	11.625	152	162762	47.326	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.660%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090221\
Data File : VV022080.D
Acq On : 02 Sep 2021 16:59
Operator : SY/MD
Sample : M3608-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKK3

Quant Time: Sep 03 01:02:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 03 00:59:14 2021
Response via : Initial Calibration

