

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090221\
 Data File : VV022083.D
 Acq On : 02 Sep 2021 18:12
 Operator : SY/MD
 Sample : M3608-10
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKK6

Quant Time: Sep 03 01:02:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 03 00:59:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	365408	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	338098	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	164670	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	103419	40.869	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.740%
7) Chloroethane-d5	1.561	69	85052	43.040	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.080%
11) 1,1-Dichloroethene-d2	2.105	63	145959	34.877	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.760%
21) 2-Butanone-d5	3.876	46	178845	103.054	ug/L	-0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.050%
24) Chloroform-d	4.349	84	206288	44.708	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.420%
26) 1,2-Dichloroethane-d4	5.034	65	128012	47.101	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.200%
32) Benzene-d6	5.050	84	434634	47.395	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.800%
36) 1,2-Dichloropropane-d6	6.072	67	135820	47.095	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.200%
41) Toluene-d8	7.317	98	397424	45.910	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.820%
43) trans-1,3-Dichloroprop...	7.625	79	59306	43.206	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.420%
47) 2-Hexanone-d5	8.088	63	117470	106.157	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.160%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	184183	49.866	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.740%
66) 1,2-Dichlorobenzene-d4	11.625	152	157555	48.670	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.340%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090221\
Data File : VV022083.D
Acq On : 02 Sep 2021 18:12
Operator : SY/MD
Sample : M3608-10
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKK6

Quant Time: Sep 03 01:02:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 03 00:59:14 2021
Response via : Initial Calibration

