

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090319\
 Data File : VV012550.D
 Acq On : 03 Sep 2019 16:22
 Operator : SY/MD
 Sample : K4607-09DL 5X
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFD65DL

Quant Time: Sep 04 02:53:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Sep 04 02:10:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	149859	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	151480	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	55810	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	56894	3.91	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.20%
7) Chloroethane-d5	1.58	69	64006	4.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.00%
11) 1,1-Dichloroethene-d2	2.13	63	108382	3.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.00%
20) 2-Butanone-d5	3.94	46	121644	43.86	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.72%
24) Chloroform-d	4.40	84	98407	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.00%
26) 1,2-Dichloroethane-d4	5.08	65	52456	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.40%
32) Benzene-d6	5.10	84	180629	4.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.60%
36) 1,2-Dichloropropane-d6	6.12	67	60021	4.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.60%
41) Toluene-d8	7.36	98	145683	3.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.60%
43) trans-1,3-Dichloropropene-	7.66	79	18647	3.81	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.20%
46) 2-Hexanone-d5	8.13	63	73137	39.87	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.74%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	39362	4.15	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	48238	4.71	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.20%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.14	96	9446	0.487 ug/L	# 1
22) cis-1,2-Dichloroethene	3.96	96	13689	0.855 ug/L	97
34) Trichloroethene	5.96	95	76483	4.438 ug/L	98
47) Tetrachloroethene	8.02	164	76943	6.179 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090319\
Data File : VV012550.D
Acq On : 03 Sep 2019 16:22
Operator : SY/MD
Sample : K4607-09DL 5X
Misc : 25mL/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFD65DL

Quant Time: Sep 04 02:53:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 04 02:10:17 2019
Response via : Initial Calibration

