

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090319\
 Data File : VV012551.D
 Acq On : 03 Sep 2019 16:48
 Operator : SY/MD
 Sample : K4607-11DL 5X
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFD67DL

Quant Time: Sep 04 03:44:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Sep 04 02:10:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	122468	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	125638	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	44204	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	62593	5.26	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.20%
7) Chloroethane-d5	1.58	69	67479	5.74	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.80%
11) 1,1-Dichloroethene-d2	2.13	63	109053	3.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.20%
20) 2-Butanone-d5	3.96	46	133910	59.08	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.16%
24) Chloroform-d	4.40	84	100771	5.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.20%
26) 1,2-Dichloroethane-d4	5.08	65	55552	5.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.20%
32) Benzene-d6	5.10	84	186788	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.60%
36) 1,2-Dichloropropane-d6	6.12	67	61781	5.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.20%
41) Toluene-d8	7.36	98	150532	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
43) trans-1,3-Dichloropropene-	7.66	79	19564	4.81	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.20%
46) 2-Hexanone-d5	8.14	63	71655	47.09	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.18%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	40987	5.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	48097	5.93	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.60%

Target Compounds					Ovalue
34) Trichloroethene	5.96	95	117082	8.191 ug/L	97
47) Tetrachloroethene	8.02	164	6783	0.657 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090319\
Data File : VV012551.D
Acq On : 03 Sep 2019 16:48
Operator : SY/MD
Sample : K4607-11DL 5X
Misc : 25mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFD67DL

Quant Time: Sep 04 03:44:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 04 02:10:17 2019
Response via : Initial Calibration

