

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090320\
 Data File : VV018149.D
 Acq On : 03 Sep 2020 11:56
 Operator : SY/MD
 Sample : VV0903WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK601

Quant Time: Sep 04 02:23:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM080619W.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 04 02:12:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	266960	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	262244	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.27	152	129663	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	76306	52.29	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	104.58%
7) Chloroethane-d5	1.58	69	74630	72.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	144.10%#
11) 1,1-Dichloroethene-d2	2.12	63	122399	48.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.76%
21) 2-Butanone-d5	3.92	46	137581	129.40	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	129.40%
24) Chloroform-d	4.38	84	168236	45.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.90%
26) 1,2-Dichloroethane-d4	5.06	65	115097	48.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.86%
32) Benzene-d6	5.08	84	334937	44.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.76%
36) 1,2-Dichloropropane-d6	6.09	67	113087	51.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.46%
41) Toluene-d8	7.34	98	292656	40.22	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.44%
43) trans-1,3-Dichloropropene-	7.64	79	52591	42.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.62%
47) 2-Hexanone-d5	8.11	63	69972	77.07	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	77.07%
56) 1,1,2,2-Tetrachloroethane-	10.24	84	145109	47.10	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.20%
66) 1,2-Dichlorobenzene-d4	11.65	152	132148	43.34	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.68%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090320\
Data File : VV018149.D
Acq On : 03 Sep 2020 11:56
Operator : SY/MD
Sample : VV0903WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK601

Quant Time: Sep 04 02:23:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM080619W.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 04 02:12:05 2020
Response via : Initial Calibration

