

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090419\
 Data File : VV012565.D
 Acq On : 04 Sep 2019 13:59
 Operator : SY/MD
 Sample : K4682-03
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0BL9

Quant Time: Sep 05 02:11:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 05 01:26:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	116770	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	122944	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	42023	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	55981	4.93	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.60%
7) Chloroethane-d5	1.58	69	58992	5.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.20%
11) 1,1-Dichloroethene-d2	2.13	63	105671	3.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.40%
20) 2-Butanone-d5	3.97	46	146946	68.00	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	136.00%#
24) Chloroform-d	4.40	84	96064	5.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.20%
26) 1,2-Dichloroethane-d4	5.08	65	53388	5.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	118.00%
32) Benzene-d6	5.10	84	172968	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.80%
36) 1,2-Dichloropropane-d6	6.12	67	56790	5.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.40%
41) Toluene-d8	7.36	98	131800	4.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.20%
43) trans-1,3-Dichloropropene-	7.66	79	17412	4.38	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.60%
46) 2-Hexanone-d5	8.14	63	79282	53.25	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.50%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	42296	5.50	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	110.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	45782	5.94	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090419\
Data File : VV012565.D
Acq On : 04 Sep 2019 13:59
Operator : SY/MD
Sample : K4682-03
Misc : 25mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0BL9

Quant Time: Sep 05 02:11:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 05 01:26:42 2019
Response via : Initial Calibration

