

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090419\
 Data File : VV012557.D
 Acq On : 04 Sep 2019 10:06
 Operator : SY/MD
 Sample : VV0904WBL01
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK48

Quant Time: Sep 05 01:39:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 05 01:26:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	131515	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	132210	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	46230	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	61530	4.82	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.40%
7) Chloroethane-d5	1.58	69	63377	5.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.40%
11) 1,1-Dichloroethene-d2	2.13	63	108378	3.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.60%
20) 2-Butanone-d5	3.97	46	126192	51.85	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.70%
24) Chloroform-d	4.40	84	99117	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.80%
26) 1,2-Dichloroethane-d4	5.08	65	54234	5.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.60%
32) Benzene-d6	5.10	84	178522	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
36) 1,2-Dichloropropane-d6	6.12	67	59662	5.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.00%
41) Toluene-d8	7.36	98	140090	4.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.20%
43) trans-1,3-Dichloropropene-	7.66	79	18356	4.29	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.80%
46) 2-Hexanone-d5	8.14	63	67960	42.44	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.88%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	39048	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	46346	5.46	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090419\
Data File : VV012557.D
Acq On : 04 Sep 2019 10:06
Operator : SY/MD
Sample : VV0904WBL01
Misc : 25mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK48

Quant Time: Sep 05 01:39:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 05 01:26:42 2019
Response via : Initial Calibration

