

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV090420\
 Data File : VV018156.D
 Acq On : 04 Sep 2020 13:33
 Operator : SY/MD
 Sample : VV0904MBL01
 Misc : 5.0g/5.0ml/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK62

Quant Time: Sep 07 02:07:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 05 00:02:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	242450	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	243466	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	119363	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	63583	34.37	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.74%
7) Chloroethane-d5	1.58	69	65467	43.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.52%
11) 1,1-Dichloroethene-d2	2.12	63	107507	30.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.76%
21) 2-Butanone-d5	3.92	46	144306	119.10	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	119.10%
24) Chloroform-d	4.38	84	162458	48.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.48%
26) 1,2-Dichloroethane-d4	5.06	65	115263	52.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.40%
32) Benzene-d6	5.08	84	312032	46.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.00%
36) 1,2-Dichloropropane-d6	6.09	67	108327	49.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.76%
41) Toluene-d8	7.34	98	269145	44.60	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.20%
43) trans-1,3-Dichloropropene-	7.64	79	52533	52.80	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	105.60%
47) 2-Hexanone-d5	8.11	63	72222	102.60	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.60%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	148027	51.82	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.64%
64) 1,2-Dichlorobenzene-d4	11.65	152	129250	55.02	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.04%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV090420\
Data File : VV018156.D
Acq On : 04 Sep 2020 13:33
Operator : SY/MD
Sample : VV0904MBL01
Misc : 5.0g/5.0ml/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK62

Quant Time: Sep 07 02:07:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM081120WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 05 00:02:28 2020
Response via : Initial Calibration

