

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090518\
 Data File : VV007397.D
 Acq On : 05 Sep 2018 21:13
 Operator : SY/MD
 Sample : J4719-10 50X
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BEGE6

Quant Time: Sep 06 02:01:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090418WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 06 01:44:11 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	486534	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	485245	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	180935	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	150101	43.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.38%
7) Chloroethane-d5	1.56	69	121201	53.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	106.48%
11) 1,1-Dichloroethene-d2	2.13	63	207531	33.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.80%
21) 2-Butanone-d5	3.96	46	226018	79.96	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	79.96%
24) Chloroform-d	4.40	84	336062	46.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.50%
26) 1,2-Dichloroethane-d4	5.09	65	214807	49.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.06%
32) Benzene-d6	5.10	84	652771	44.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.26%
36) 1,2-Dichloropropane-d6	6.13	67	226855	46.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.58%
41) Toluene-d8	7.36	98	519304	41.25	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.50%
43) trans-1,3-Dichloropropene-	7.67	79	88572	42.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.86%
47) 2-Hexanone-d5	8.14	63	164881	85.82	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.82%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	318574	47.59	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.18%
64) 1,2-Dichlorobenzene-d4	11.68	152	202618	52.10	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090518\
Data File : VV007397.D
Acq On : 05 Sep 2018 21:13
Operator : SY/MD
Sample : J4719-10 50X
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BEGE6

Quant Time: Sep 06 02:01:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090418WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 06 01:44:11 2018
Response via : Initial Calibration

