

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090524\
 Data File : VV037206.D
 Acq On : 05 Sep 2024 09:33
 Operator : SY/MD
 Sample : VV0905SBL01
 Misc : 5.00g/10mL/MSVOA_V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK399

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 09/06/2024
 Supervised By :Mahesh Dadoda 09/06/2024

Quant Time: Sep 06 01:31:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM090424SMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 06 01:31:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	561946	25.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	537088	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	251704	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	201756	20.044	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.160%
7) Chloroethane-d5	1.532	69	163669	20.947	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.800%
11) 1,1-Dichloroethene-d2	2.060	65	94004	20.762	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	83.040%
21) 2-Butanone-d5	3.838	46	93337m	42.876	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.760%
24) Chloroform-d	4.243	84	375225	21.700	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.800%
26) 1,2-Dichloroethane-d4	4.940	65	201543	22.979	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.920%
32) Benzene-d6	4.953	84	772512	22.762	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	91.040%
36) 1,2-Dichloropropane-d6	5.985	67	243216	23.133	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.520%
41) Toluene-d8	7.239	98	685851	22.346	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.400%
43) trans-1,3-Dichloroprop...	7.551	79	88152	21.781	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.120%
47) 2-Hexanone-d5	8.034	63	54460	37.997	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	76.000%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	178415	22.077	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	88.320%
66) 1,2-Dichlorobenzene-d4	11.554	152	235246	23.943	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090524\
Data File : VV037206.D
Acq On : 05 Sep 2024 09:33
Operator : SY/MD
Sample : VV0905SBL01
Misc : 5.00g/10mL/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK399

Quant Time: Sep 06 01:31:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM090424SMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 06 01:31:19 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 09/06/2024
Supervised By :Mahesh Dadoda 09/06/2024

