

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090618\
 Data File : VV007428.D
 Acq On : 06 Sep 2018 18:42
 Operator : SY/MD
 Sample : J4719-08
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BEGE4

Quant Time: Sep 07 02:12:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090418WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 07 01:29:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	517465	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	511025	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	194305	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	163796	44.32	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.64%
7) Chloroethane-d5	1.56	69	126382	52.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	104.40%
11) 1,1-Dichloroethene-d2	2.13	63	217379	33.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.78%
21) 2-Butanone-d5	3.96	46	200155	66.57	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	66.57%
24) Chloroform-d	4.41	84	318057	41.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.20%
26) 1,2-Dichloroethane-d4	5.09	65	201771	43.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.50%
32) Benzene-d6	5.10	84	644540	41.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.70%
36) 1,2-Dichloropropane-d6	6.13	67	214705	41.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.20%
41) Toluene-d8	7.36	98	521298	39.32	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.64%#
43) trans-1,3-Dichloropropene-	7.67	79	85600	38.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.88%
47) 2-Hexanone-d5	8.14	63	136379	67.40	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	67.40%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	286213	40.60	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	194294	46.53	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.06%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chloromethane	1.26	50	78871	18.10	ug/L	95
6) Bromomethane	1.51	94	6466	3.04	ug/L	100
13) Acetone	2.20	43	104274	58.24	ug/L	98
22) 2-Butanone	4.05	43	13727	4.58	ug/L	99
29) Cyclohexane	4.72	56	10064	1.78	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090618\
Data File : VV007428.D
Acq On : 06 Sep 2018 18:42
Operator : SY/MD
Sample : J4719-08
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BEGE4

Quant Time: Sep 07 02:12:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090418WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 07 01:29:25 2018
Response via : Initial Calibration

