

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090619\
 Data File : VV012611.D
 Acq On : 06 Sep 2019 21:20
 Operator : SY/MD
 Sample : K4703-10
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFD93

Quant Time: Sep 07 02:55:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090519WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 07 01:06:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	143716	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	146890	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	47695	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	64133	4.77	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.40%
7) Chloroethane-d5	1.58	69	64217	5.27	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.40%
11) 1,1-Dichloroethene-d2	2.13	63	105676	3.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.00%
20) 2-Butanone-d5	3.97	46	141335	53.92	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.84%
24) Chloroform-d	4.40	84	100616	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
26) 1,2-Dichloroethane-d4	5.08	65	56572	5.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.00%
32) Benzene-d6	5.10	84	181960	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
36) 1,2-Dichloropropane-d6	6.12	67	61691	5.17	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.40%
41) Toluene-d8	7.36	98	135182	4.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.20%
43) trans-1,3-Dichloropropene-	7.67	79	17196	4.10	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.00%
46) 2-Hexanone-d5	8.14	63	70747	44.20	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.40%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	41237	4.88	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	43951	5.52	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.40%

Target Compounds					Ovalue
34) Trichloroethene	5.96	95	11166	0.754 ug/L	94
47) Tetrachloroethene	8.02	164	74196	6.871 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090619\
Data File : VV012611.D
Acq On : 06 Sep 2019 21:20
Operator : SY/MD
Sample : K4703-10
Misc : 25mL/MSVOA V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFD93

Quant Time: Sep 07 02:55:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090519WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 07 01:06:07 2019
Response via : Initial Calibration

