

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV090619\
 Data File : VV012602.D
 Acq On : 06 Sep 2019 17:22
 Operator : SY/MD
 Sample : K4682-05
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0BM0

Quant Time: Sep 07 01:31:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090519WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 07 01:06:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	146077	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	150076	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	52752	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	54763	4.01	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.20%
7) Chloroethane-d5	1.58	69	60503	4.89	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.80%
11) 1,1-Dichloroethene-d2	2.13	63	93307	2.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.20%#
20) 2-Butanone-d5	3.93	46	138392	51.94	ug/L	-0.04
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.88%
24) Chloroform-d	4.40	84	98911	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.20%
26) 1,2-Dichloroethane-d4	5.08	65	54130	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
32) Benzene-d6	5.10	84	175967	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
36) 1,2-Dichloropropane-d6	6.12	67	60507	4.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.40%
41) Toluene-d8	7.36	98	135557	4.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.80%
43) trans-1,3-Dichloropropene-	7.66	79	18585	4.34	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.80%
46) 2-Hexanone-d5	8.13	63	67158	41.06	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.12%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	40051	4.64	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	44335	5.03	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV090619\
Data File : VV012602.D
Acq On : 06 Sep 2019 17:22
Operator : SY/MD
Sample : K4682-05
Misc : 25mL/MSVOA V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0BM0

Quant Time: Sep 07 01:31:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR090519WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 07 01:06:07 2019
Response via : Initial Calibration

