

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090622\
 Data File : VV027750.D
 Acq On : 06 Sep 2022 18:51
 Operator : SY/MD
 Sample : VV0906WBL02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK226

Quant Time: Sep 07 06:13:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 07 06:13:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.629	114	150928	5.000	ug/L	0.04
28) Chlorobenzene-d5	8.866	117	133336	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	59019	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	50563	5.082	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.600%
7) Chloroethane-d5	1.581	69	43365	5.138	ug/L	0.02
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.800%
11) 1,1-Dichloroethene-d2	2.137	63	82457	3.906	ug/L	0.04
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.200%
20) 2-Butanone-d5	3.995	46	133737	60.657	ug/L	0.05
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.320%
24) Chloroform-d	4.378	84	111262	5.462	ug/L	0.03
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.200%
26) 1,2-Dichloroethane-d4	5.050	65	62470	5.759	ug/L	0.03
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.200%
32) Benzene-d6	5.056	84	218185	5.640	ug/L	0.03
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.800%
36) 1,2-Dichloropropane-d6	6.088	67	55204	4.874	ug/L	0.04
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.400%
41) Toluene-d8	7.342	98	145600	4.476	ug/L	0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
43) trans-1,3-Dichloroprop...	7.651	79	17038	4.484	ug/L	0.02
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.600%
46) 2-Hexanone-d5	8.121	63	64515	45.681	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.360%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	36886	5.072	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	50524	5.211	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090622\
Data File : VV027750.D
Acq On : 06 Sep 2022 18:51
Operator : SY/MD
Sample : VV0906WBL02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK226

Quant Time: Sep 07 06:13:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 07 06:13:02 2022
Response via : Initial Calibration

