

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090721\
 Data File : VV022104.D
 Acq On : 07 Sep 2021 15:48
 Operator : SY/MD
 Sample : M3608-17
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKL3

Quant Time: Sep 08 01:17:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 08 01:14:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	344969	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	322007	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	145439	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	79688	33.357	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	66.720%
7) Chloroethane-d5	1.564	69	68814	36.886	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	73.780%
11) 1,1-Dichloroethene-d2	2.105	63	120734	30.559	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.120%
21) 2-Butanone-d5	3.892	46	166731	101.765	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.770%
24) Chloroform-d	4.352	84	184596	42.377	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.760%
26) 1,2-Dichloroethane-d4	5.034	65	117146	45.657	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.320%
32) Benzene-d6	5.050	84	374224	42.847	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.700%
36) 1,2-Dichloropropane-d6	6.072	67	122156	44.474	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.940%
41) Toluene-d8	7.317	98	327333	39.703	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.400%#
43) trans-1,3-Dichloroprop...	7.625	79	51771	39.601	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.200%
47) 2-Hexanone-d5	8.092	63	103122	97.848	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.850%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	164829	46.856	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.720%
66) 1,2-Dichlorobenzene-d4	11.625	152	130541	45.657	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090721\
Data File : VV022104.D
Acq On : 07 Sep 2021 15:48
Operator : SY/MD
Sample : M3608-17
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKL3

Quant Time: Sep 08 01:17:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 08 01:14:49 2021
Response via : Initial Calibration

