

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090721\
 Data File : VV022111.D
 Acq On : 07 Sep 2021 19:25
 Operator : SY/MD
 Sample : M3651-05
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKM1

Quant Time: Sep 08 01:19:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 08 01:14:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	341636	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	314333	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	159256	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	81697	34.531	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.060%
7) Chloroethane-d5	1.565	69	71351	38.619	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	77.240%
11) 1,1-Dichloroethene-d2	2.105	63	119784	30.614	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.220%
21) 2-Butanone-d5	3.896	46	164785	101.559	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.560%
24) Chloroform-d	4.352	84	180644	41.875	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.740%
26) 1,2-Dichloroethane-d4	5.034	65	114757	45.162	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.320%
32) Benzene-d6	5.053	84	364366	42.737	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.480%
36) 1,2-Dichloropropane-d6	6.072	67	118902	44.346	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.700%
41) Toluene-d8	7.317	98	326593	40.580	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.160%
43) trans-1,3-Dichloroprop...	7.625	79	48610	38.091	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.180%
47) 2-Hexanone-d5	8.092	63	108190	105.163	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.160%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	170273	49.585	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.180%
66) 1,2-Dichlorobenzene-d4	11.625	152	142917	45.649	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.300%
Target Compounds						Qvalue
13) Acetone	2.188	43	6529	4.803	ug/L	# 89
33) Benzene	5.105	78	12344	1.350	ug/L	100
42) Toluene	7.394	91	48105	4.874	ug/L	98
52) Ethylbenzene	9.014	91	116642	10.917	ug/L	99
53) m,p-Xylene	9.140	106	20581	4.956	ug/L	93
54) o-Xylene	9.548	106	13414	3.264	ug/L	100
55) Styrene	9.571	104	6879	0.981	ug/L	97
63) 1,2,4-Trimethylbenzene	10.918	105	16231	1.901	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090721\
Data File : VV022111.D
Acq On : 07 Sep 2021 19:25
Operator : SY/MD
Sample : M3651-05
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKM1

Quant Time: Sep 08 01:19:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 08 01:14:49 2021
Response via : Initial Calibration

