

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090721\
 Data File : VV022112.D
 Acq On : 07 Sep 2021 19:49
 Operator : SY/MD
 Sample : M3651-10
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKM4

Quant Time: Sep 08 01:19:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 08 01:14:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	354133	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	328936	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	151021	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	80319	32.751	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	65.500%
7) Chloroethane-d5	1.565	69	69607	36.345	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	72.700%
11) 1,1-Dichloroethene-d2	2.105	63	118886	29.312	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.620%#
21) 2-Butanone-d5	3.892	46	170870	101.593	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.590%
24) Chloroform-d	4.349	84	182371	40.783	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.560%
26) 1,2-Dichloroethane-d4	5.034	65	114897	43.622	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.240%
32) Benzene-d6	5.050	84	365878	41.009	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.020%
36) 1,2-Dichloropropane-d6	6.072	67	120604	42.984	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.960%
41) Toluene-d8	7.317	98	322486	38.291	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	76.580%#
43) trans-1,3-Dichloroprop...	7.625	79	49692	37.210	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.420%
47) 2-Hexanone-d5	8.092	63	105801	98.275	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.280%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	164660	45.822	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.640%
66) 1,2-Dichlorobenzene-d4	11.625	152	131376	44.251	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.500%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090721\
Data File : VV022112.D
Acq On : 07 Sep 2021 19:49
Operator : SY/MD
Sample : M3651-10
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKM4

Quant Time: Sep 08 01:19:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 08 01:14:49 2021
Response via : Initial Calibration

