

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090721\
 Data File : VV022115.D
 Acq On : 07 Sep 2021 21:01
 Operator : SY/MD
 Sample : M3651-13
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKM7

Quant Time: Sep 08 01:20:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 08 01:14:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	329646	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	310997	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	138853	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	74681	32.714	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	65.420%
7) Chloroethane-d5	1.564	69	66742	37.438	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	74.880%
11) 1,1-Dichloroethene-d2	2.105	63	112137	29.702	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.400%#
21) 2-Butanone-d5	3.905	46	164952	105.360	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.360%
24) Chloroform-d	4.352	84	177277	42.589	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.180%
26) 1,2-Dichloroethane-d4	5.037	65	111997	45.679	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.360%
32) Benzene-d6	5.050	84	352694	41.812	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.620%
36) 1,2-Dichloropropane-d6	6.072	67	115024	43.360	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.720%
41) Toluene-d8	7.320	98	310851	39.038	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.080%#
43) trans-1,3-Dichloroprop...	7.625	79	46084	36.499	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.000%
47) 2-Hexanone-d5	8.091	63	97942	96.223	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.220%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	156937	46.192	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.380%
66) 1,2-Dichlorobenzene-d4	11.625	152	122620	44.921	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090721\
Data File : VV022115.D
Acq On : 07 Sep 2021 21:01
Operator : SY/MD
Sample : M3651-13
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKM7

Quant Time: Sep 08 01:20:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 08 01:14:49 2021
Response via : Initial Calibration

