

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090721\
 Data File : VV022108.D
 Acq On : 07 Sep 2021 18:12
 Operator : SY/MD
 Sample : M3651-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKL7

Quant Time: Sep 08 01:18:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 08 01:14:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	337688	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	312648	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	141467	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	82686	35.358	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	70.720%
7) Chloroethane-d5	1.565	69	72699	39.808	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.620%
11) 1,1-Dichloroethene-d2	2.105	63	123598	31.958	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.920%
21) 2-Butanone-d5	3.899	46	171430	106.890	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.890%
24) Chloroform-d	4.352	84	187206	43.903	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.800%
26) 1,2-Dichloroethane-d4	5.034	65	116896	46.542	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.080%
32) Benzene-d6	5.050	84	376864	44.441	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.880%
36) 1,2-Dichloropropane-d6	6.069	67	122101	45.784	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.560%
41) Toluene-d8	7.317	98	329862	41.207	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.420%
43) trans-1,3-Dichloroprop...	7.625	79	52097	41.043	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.080%
47) 2-Hexanone-d5	8.092	63	107183	104.745	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.750%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	163716	47.933	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.860%
66) 1,2-Dichlorobenzene-d4	11.625	152	128805	46.315	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090721\
Data File : VV022108.D
Acq On : 07 Sep 2021 18:12
Operator : SY/MD
Sample : M3651-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKL7

Quant Time: Sep 08 01:18:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 08 01:14:49 2021
Response via : Initial Calibration

