

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW090721\  
 Data File : VW022112.D  
 Acq On : 07 Sep 2021 19:49  
 Operator : SY/MD  
 Sample : M3651-10  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 DBKM4

Quant Time: Sep 08 01:19:40 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVLM082721WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Wed Sep 08 01:14:49 2021  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	354133	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	328936	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	151021	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	80319	32.751	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery =	65.500%		
7) Chloroethane-d5	1.565	69	69607	36.345	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery =	72.700%		
11) 1,1-Dichloroethene-d2	2.105	63	118886	29.312	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery =	58.620%#		
21) 2-Butanone-d5	3.892	46	170870	101.593	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery =	101.590%		
24) Chloroform-d	4.349	84	182371	40.783	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	81.560%		
26) 1,2-Dichloroethane-d4	5.034	65	114897	43.622	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	87.240%		
32) Benzene-d6	5.050	84	365878	41.009	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery =	82.020%		
36) 1,2-Dichloropropane-d6	6.072	67	120604	42.984	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery =	85.960%		
41) Toluene-d8	7.317	98	322486	38.291	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery =	76.580%#		
43) trans-1,3-Dichloroprop...	7.625	79	49692	37.210	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery =	74.420%		
47) 2-Hexanone-d5	8.092	63	105801	98.275	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery =	98.280%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	164660	45.822	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery =	91.640%		
66) 1,2-Dichlorobenzene-d4	11.625	152	131376	44.251	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery =	88.500%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV090721\  
 Data File : VV022112.D  
 Acq On : 07 Sep 2021 19:49  
 Operator : SY/MD  
 Sample : M3651-10  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampled :  
 DBKM4

Quant Time: Sep 08 01:19:40 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVLM082721WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Wed Sep 08 01:14:49 2021  
 Response via : Initial Calibration

