

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090722\
 Data File : VV027773.D
 Acq On : 07 Sep 2022 11:12
 Operator : SY/MD
 Sample : VV0907WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBK227

Quant Time: Sep 08 07:54:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 07 06:13:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	138284	5.000	ug/L	0.02
28) Chlorobenzene-d5	8.847	117	136268	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	63786	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	47028	5.159	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	103.200%	
7) Chloroethane-d5	1.564	69	41354	5.348	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	107.000%	
11) 1,1-Dichloroethene-d2	2.105	63	73570	3.804	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	76.000%	
20) 2-Butanone-d5	3.902	46	80199	39.700	ug/L	-0.04
Spiked Amount	50.000	Range 40 - 130	Recovery	=	79.400%	
24) Chloroform-d	4.339	84	96291	5.159	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	103.200%	
26) 1,2-Dichloroethane-d4	5.027	65	46602	4.689	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.800%	
32) Benzene-d6	5.043	84	169550	4.289	ug/L	0.01
Spiked Amount	5.000	Range 70 - 125	Recovery	=	85.800%	
36) 1,2-Dichloropropane-d6	6.063	67	49073	4.240	ug/L	0.01
Spiked Amount	5.000	Range 60 - 140	Recovery	=	84.800%	
41) Toluene-d8	7.310	98	143284	4.310	ug/L	-0.01
Spiked Amount	5.000	Range 70 - 130	Recovery	=	86.200%	
43) trans-1,3-Dichloroprop...	7.619	79	17851	4.597	ug/L	-0.02
Spiked Amount	5.000	Range 55 - 130	Recovery	=	92.000%	
46) 2-Hexanone-d5	8.085	63	59985	41.559	ug/L	-0.03
Spiked Amount	50.000	Range 45 - 130	Recovery	=	83.120%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	31536	4.243	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	84.800%	
66) 1,2-Dichlorobenzene-d4	11.622	152	50617	4.831	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	96.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090722\
 Data File : VV027773.D
 Acq On : 07 Sep 2022 11:12
 Operator : SY/MD
 Sample : VV0907WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK227

Quant Time: Sep 08 07:54:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 07 06:13:02 2022
 Response via : Initial Calibration

