

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090821\
 Data File : VV022123.D
 Acq On : 08 Sep 2021 12:46
 Operator : SY/MD
 Sample : M3651-15
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBKM9

Quant Time: Sep 09 08:28:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 09 08:27:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	321744	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	299988	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	134683	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	62395	28.003	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	56.000%#
7) Chloroethane-d5	1.558	69	62434	35.882	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	71.760%
11) 1,1-Dichloroethene-d2	2.101	63	98664	26.775	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	53.560%#
21) 2-Butanone-d5	3.873	46	157807	103.271	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.270%
24) Chloroform-d	4.349	84	168306	41.427	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.860%
26) 1,2-Dichloroethane-d4	5.034	65	108095	45.171	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.340%
32) Benzene-d6	5.050	84	337372	41.463	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.920%
36) 1,2-Dichloropropane-d6	6.072	67	113004	44.162	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.320%
41) Toluene-d8	7.317	98	295354	38.453	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	76.900%#
43) trans-1,3-Dichloroprop...	7.625	79	45515	37.371	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.740%
47) 2-Hexanone-d5	8.088	63	98846	100.675	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.670%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	161570	49.301	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	120951	45.681	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090821\
Data File : VV022123.D
Acq On : 08 Sep 2021 12:46
Operator : SY/MD
Sample : M3651-15
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBKM9

Quant Time: Sep 09 08:28:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 09 08:27:48 2021
Response via : Initial Calibration

