

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090822\
 Data File : VV027784.D
 Acq On : 08 Sep 2022 12:44
 Operator : SY/MD
 Sample : VSTD0.108
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.1208

Quant Time: Sep 09 01:46:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM090822.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 09 01:45:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Difluorobenzene	5.615	114	4779	0.500 ug/L	0.00
5) Chlorobenzene-d5	8.848	117	4387	0.500 ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.244	152	3007	0.500 ug/L	0.00
System Monitoring Compounds					
2) Vinyl Chloride-d3	1.310	65	224	0.065 ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	165	0.078 ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	113	0.049 ug/L	0.00
8) Toluene-d8	7.317	98	571	0.060 ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	96	0.053 ug/L	0.00
Target Compounds					
					Qvalue
3) Vinyl chloride	1.310	62	245	0.082 ug/L #	72
6) Trichloroethene	5.925	95	228	0.075 ug/L	86
9) 1,2-Dibromoethane	8.368	107	145	0.078 ug/L #	81
12) 1,2,3-Trichloropropane	10.267	75	76	0.059 ug/L #	86
13) 1,2-Dibromo-3-chloropr...	12.409	75	19	0.079 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090822\
Data File : VV027784.D
Acq On : 08 Sep 2022 12:44
Operator : SY/MD
Sample : VSTD0.108
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.1208

Quant Time: Sep 09 01:46:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM090822.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 09 01:45:15 2022
Response via : Initial Calibration

