

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090822\
 Data File : VV027788.D
 Acq On : 08 Sep 2022 14:22
 Operator : SY/MD
 Sample : VIBLK201
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK201

Quant Time: Sep 09 02:57:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM090822.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 09 02:55:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	4477	0.500	ug/L	0.02
5) Chlorobenzene-d5	8.864	117	3988	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	2717	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	1557	0.561	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 112.000%	
4) 1,2-Dichloroethane-d4	5.042	65	1148	0.550	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 110.000%	
7) 1,2-Dichloropropane-d6	6.092	67	739	0.556	ug/L	0.02
Spiked Amount	0.500	Range	60 - 140	Recovery	= 112.000%	
8) Toluene-d8	7.332	98	4277	0.582	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	= 116.000%	
10) 1,1,2,2-Tetrachloroeth...	10.216	84	655	0.576	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 116.000%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090822\
Data File : VV027788.D
Acq On : 08 Sep 2022 14:22
Operator : SY/MD
Sample : VIBLK201
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK201

Quant Time: Sep 09 02:57:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM090822.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 09 02:55:15 2022
Response via : Initial Calibration

