

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090823\
 Data File : VV031978.D
 Acq On : 08 Sep 2023 12:18
 Operator : SY/MD
 Sample : 04277-02 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JYCA1

Quant Time: Sep 09 05:34:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 09 05:33:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	286069	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	275807	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	140225	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	110113	61.505	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 123.000%	
7) Chloroethane-d5	1.529	69	89344	56.928	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 113.860%	
11) 1,1-Dichloroethene-d2	2.060	65	46583	55.567	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 111.140%	
21) 2-Butanone-d5	3.789	46	181454	128.866	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 128.870%	
24) Chloroform-d	4.252	84	205909	52.619	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 105.240%	
26) 1,2-Dichloroethane-d4	4.944	65	136423	52.533	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 105.060%	
32) Benzene-d6	4.963	84	432935	57.526	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 115.060%	
36) 1,2-Dichloropropane-d6	5.992	67	140910	58.171	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 116.340%	
41) Toluene-d8	7.246	98	397300	57.294	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 114.580%	
43) trans-1,3-Dichloroprop...	7.558	79	70186	55.902	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 111.800%	
47) 2-Hexanone-d5	8.027	63	123993	167.006	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 167.010%#	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	183304	58.559	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 117.120%	
66) 1,2-Dichlorobenzene-d4	11.564	152	145941	57.092	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 114.180%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090823\
Data File : VV031978.D
Acq On : 08 Sep 2023 12:18
Operator : SY/MD
Sample : 04277-02 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JYCA1

Quant Time: Sep 09 05:34:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 09 05:33:49 2023
Response via : Initial Calibration

