

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW090823\
 Data File : VW031981.D
 Acq On : 08 Sep 2023 13:29
 Operator : SY/MD
 Sample : 04280-03 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JLAD2

Quant Time: Sep 09 05:35:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 09 05:33:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	280093	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	269123	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	139046	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	82962	47.328	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.660%
7) Chloroethane-d5	1.529	69	66415	43.221	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.440%
11) 1,1-Dichloroethene-d2	2.060	65	35360	43.080	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.160%
21) 2-Butanone-d5	3.793	46	137268	99.566	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.570%
24) Chloroform-d	4.252	84	154756	40.391	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.780%
26) 1,2-Dichloroethane-d4	4.947	65	102508	40.315	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.640%
32) Benzene-d6	4.963	84	325558	44.332	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.660%
36) 1,2-Dichloropropane-d6	5.995	67	105804	44.763	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.520%
41) Toluene-d8	7.249	98	298683	44.142	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.280%
43) trans-1,3-Dichloroprop...	7.558	79	52194	42.604	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.200%
47) 2-Hexanone-d5	8.030	63	94934	131.042	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	131.040%#
56) 1,1,2,2-Tetrachloroeth...	10.159	84	140116	45.874	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.740%
66) 1,2-Dichlorobenzene-d4	11.564	152	109806	43.320	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090823\
Data File : VV031981.D
Acq On : 08 Sep 2023 13:29
Operator : SY/MD
Sample : 04280-03 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JLAD2

Quant Time: Sep 09 05:35:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 09 05:33:49 2023
Response via : Initial Calibration

